

**ANALOX 100HYP MKIII
OXYGEN MONITOR**



ON/OFF



SENSOR



CALIBRATE
±

Range: 0.0 to 1.99 Bar Partial Pressure
Sensor Life: 3 years at 0.21 Bar ppO_2
Response time: Less than 6 secs to T90
Battery: 9 volt Alkaline PP3, MN1604 etc.
Battery Life: Approximately 800 hours

OPERATION: The display indicates the partial pressure of Oxygen at the Sensor.

CALIBRATION: In fresh air at normal Atmospheric Pressure, adjust the trimmer on the front panel, until the display reads 0.21 - No zero adjustment necessary.

CAUTION: The monitor should only be subjected to pressure changes at a rate suitable for human beings. It may be damaged if subjected to rapid rates of compression or decompression. It should NOT be passed through chamber medical locks.

Analox Sensor Technology Ltd
Unit 15, Ellerbeck Court, Stokesley Business Park
Stokesley, Middlesbrough, Cleveland TS9 5PT.
www.analox.net